

# Journal Pre-proof

A converted spray device for topical pharyngeal anesthesia in awake intubation

Xue Li, Hai-yang Yu, Xiao Hu



PII: S0104-0014(21)00354-7

DOI: <https://doi.org/10.1016/j.bjane.2021.09.005>

Reference: BJANE 744282

To appear in: *Brazilian Journal of Anesthesiology (English edition)*

Received Date: 27 April 2021

Please cite this article as: Li X, Yu H-yang, Hu X, A converted spray device for topical pharyngeal anesthesia in awake intubation, *Brazilian Journal of Anesthesiology (English edition)* (2021), doi: <https://doi.org/10.1016/j.bjane.2021.09.005>

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2020 Published by Elsevier.

**BJAN-D-21-00181 - Letter to the Editor**

**A converted spray device for topical pharyngeal anesthesia in awake intubation**

**Xue Li\*, Hai-yang Yu, Xiao Hu**

Peking University First Hospital, Department of Anesthesiology, Beijing, China

**\* Corresponding author.**

E-mails: lixuepku@163.com; 3999165@163.com\_(X. Li).

ORCID ID: 0000-0003-3437-0870

Received 27 April 2021; accepted 11 September 2021

*Dear editor,*

Ankylosing Spondylitis (AS) is a common inflammatory arthritis with progressive reduced spinal mobility.[1] Difficult airway exists in those patients with restricted neck movement, in which scenario awake intubation before anesthesia induction is suggested. Adequate topical anesthesia of oropharyngeal/nasopharyngeal and subglottic tracheal mucosa is required for airway instrumentation and patient comfort during this procedure.

Traditional device for topical pharyngeal anesthesia, like laryngotracheal topical anesthesia kit, is widespread but its effectiveness is not always ideal. Other techniques are soon proposed, for example, both ultrasonic nebulization and spray-as-you-go techniques provide acceptable conditions for awake intubation.[2] However, ultrasonic nebulizer is time consuming (usually needs 10–15 min) and spray-as-you-go technique is perceived difficult. Ultrasound-guided superior laryngeal nerve block is a hot topic, but it is invasive and required bilateral block.[3] Here we reported our successful experience for awake intubation through a new spray device easily converted from a spray bottle and oxygen delivery device. This converted spray device was consisted of

three main parts, i.e., a glass bottle with a special spray nozzle, a simple humidifier, and a pipe for oxygen delivery (Fig. 1).

From October 8, 2020 to November 30, 2020, 5 patients aged from 29–49 years, diagnosed with AS undergoing spinal deformity surgery received the awake intubation. After intravenous cannula insertion, patients were connected to routine monitors and received supplemental oxygen ( $5 \text{ L}\cdot\text{min}^{-1}$ ) via ventilation mask. Intravenous sedation consisted of  $0.8\text{--}1 \mu\text{g}\cdot\text{kg}^{-1}$  dexmedetomidine with remifentanil infusion (target effect-site concentration  $0.5\text{--}1.5 \text{ ng}\cdot\text{mL}^{-1}$ ) to keep the Modified Observer's Assessment of Alertness/Sedation (MOAA/S) score between 3 to 5.[4]

About 10 min after administering dexmedetomidine, we used this converted spray device driven by oxygen ( $5 \text{ L}\cdot\text{min}^{-1}$ ) to anesthetize pharyngeal mucosa (Fig. 1). We firstly instructed the patient to open their mouth loudly, stick out their tongue as far as possible and sprayed 1% tetracaine solution into oropharynx for 3–5 seconds, then we sprayed the tetracaine solution into nasopharynx when they took a deep breath for 2–3 seconds bilaterally as well. Two minutes later, we repeated these steps. Another 1 min later, we used a 5-mL syringe for thyrocricocentesis after identifying the cricothyroid membrane by landmark palpation and sterilizing the local area, 3 mL 2% lidocaine was given for subglottic tracheal mucosa anesthesia once the tip of needle located in the trachea, as signaled by a sudden loss of resistance and freely air withdrawing.

We performed awake intubation by flexible fiberscope with the endotracheal tube in-situ through the most patent nostril. After passing through the choana and identifying the epiglottis and glottis, the fiberscope was moved slowly into the trachea until the tracheal carina was visualized. Then the endotracheal tube was railroaded through the flexible fiberscope into the trachea. After reconfirming the position of the tube by mainstream capnograph, the cuff was inflated. During the procedure, glottis opening was fine when the fiberscope advanced into the glottis and trachea, and no intolerance symptoms such as cough, vomit, movement were found. Besides, MAP, HR and  $\text{SPO}_2$  were all stable. Patients were all satisfied about the procedure. General anesthesia was induced by propofol, remifentanil and rocuronium bromide. The anesthesia maintenance and surgery of those patients were all uneventful.

This converted spray device for topical pharyngeal anesthesia has several advantages. First, it is noninvasive and is more acceptable to patient; Second, it is easily converted from routine clinical devices and is very simple to use; Third, the consumed tetracaine is very small (median dose of these 5 patients is 35 mg), thus the risk of local

anesthetic systemic toxicity is relatively low; Last and most important, it is more timesaving. Generally speaking, it only takes us 3–5 min to get sufficient topical anesthesia both for oropharynx and nasopharynx. Thus, it is a promising device for awake intubation in clinical practice.

### **Conflicts of interest**

The authors declare no conflicts of interest.

### **References**

- [1] Taurog JD, Chhabra A, Colbert RA. Ankylosing Spondylitis and Axial Spondyloarthritis. *N Engl J Med.* 2016;374:2563-74.
- [2] Dhasmana S, Singh V, Pal US. Nebulisation Versus Spray-as-You-Go Airway Topical Anaesthesia in Patients with Temporomandibular Joint Ankylosis using 2 % Lignocaine. *J Maxillofac Oral Surg.* 2015;14:398-402.
- [3] Stopar-Pintaric T, Vlassakov K, Azman J, et al. The thyrohyoid membrane as a target for ultrasonography-guided block of the internal branch of the superior laryngeal nerve. *J Clin Anesth.* 2015;27:548-52.
- [4] Chernik DA, Gillings D, Laine H, et al. Validity and reliability of the Observer's Assessment of Alertness/Sedation Scale: study with intravenous midazolam. *J Clin Psychopharmacol.* 1990;10:244-51.

**Figure 1** - Converted spray device and beautiful sprayed solution.

